



CONFORMITY CERTIFICATE OF EXPLOSION-PROOF

Certificate No.: CCRI 25.2650X

Applicant Company: FTI TECHNOLOGY INC.

Address: 1013 Centre Road, Suite 403S Wilmington, DE 19805 USA

Name of Product: pressure transmitter

Type of Product: PB330Series

Marking: Ex ia II C T4 Ga

Technical Documents: Q/FTI 001-2025

Drawing No.: PB330-000

Note(s):

- 1.Rated voltage: 24VDC, use an associated device that matches its intrinsic safety parameters.
Intrinsic safety parameters: Ui:DC28V, Ii:93mA, Pi:0.65W, Ci:10nF, Li:0mH.
- 2.IP code: IP67.
- 3.The certificate suffix "X" indicates that the device is used in Zone 0 and should be protected from ignition hazards caused by impact or friction.
- 4.Association inspection report number: 2025550341.

By verifying the drawings and technical documents and checking samples, the product complies with the following standards currently valid in P.R.China:

GB/T 3836.1-2021, GB/T 3836.4-2021, GB/T 4208-2017

Date of Issue: 5. 16. 2025

Valid till: 5. 15. 2030

Authorize:



CCRI(Beijing)TestingTechnologyCo.,Ltd.

(Beijing Test Center of Mine Working Equipment and Mining Oils for National Safe Production)

Note: This certificate is only valid for the products that are in accord with samples tested and verified.

Address: Building 6, No.5, Yulong Street, Caiyu Town, Beijing Economic and Technological Development Area(Daxing), Beijing Tel/fax: 010-89268018

The query url of Explosion-proof Certificate: <http://jc.ccrtec.com.cn>